

CERTIFICATE OF TESTING No:6648/10

Issue 1

Customer: Bioquell UK Ltd
52 Royce Close
Portway Industrial Estate
Andover, Hampshire
SP10 3TS

Customer Reference No: STD229301

TRaC ID No: 6959

Equipment Tested

EUT: Clarus Model No: Port Serial No: 2008099962

Test Specifications

BS EN 61326-1:2006

testing regulatory and compliance

Date of Test: 3rd – 8th December 2009 Test Location: TRaC Malvern

Comments: Emissions: Class A
Immunity: Industrial Locations

Test Results: Compliant – See table of tests on page 2.

This is to certify that the above referenced equipment has been tested in accordance with the requirements of the contract/order/specification as detailed above using test equipment calibrated to traceable National Standards. This Certificate applies only to the product(s) detailed and is not indicative of the qualities of identical or similar products.

Approved By:



S Youngman (General Manager)

Date of Issue: 27th January 2010

Page 1 of 2

The results herein relate only to the sample tested. Full results are contained in the relevant works order file.

RINGWOOD

Holly Grove Farm, Verwood Road, Ashley, Ringwood, Hants, BH24 2DB, UK.

T +44 (0)1425 479979 F +44 (0)1425 480637 E test@tracglobal.com

www.tracglobal.com

Tests Performed

The following tests were performed on the sample detailed on page 1 of this certificate:

Test	Basic Standard	Frequency Range	Level/Limit	Result
Radiated Emissions	EN 55011	30MHz-1GHz	Class A	Pass
Conducted Emissions		150kHz-30MHz		
Electrostatic Discharge	EN 61000-4-2	N/A	±4kV Contact Discharge ±8kV Air Discharge ±4kV Coupling Planes	Pass
Radiated RF Immunity	EN 61000-4-3	80MHz-1GHz	10V/m 80% AM 1kHz	Pass
		1.4-2GHz	3V/m 80% AM 1kHz	Pass
		2-2.7GHz	1V/m 80% AM 1kHz	Pass
EFT	EN 61000-4-4	N/A	Power Supply ±500V, ±1kV & ±2kV	Pass
Surges	EN 61000-4-5	N/A	Power Supply ±500V ±1kV & ±2kV CM ±250V ±500V & ±1kV DM	Pass
Conducted RF Immunity	EN 61000-4-6	150kHz-80MHz	3V r.m.s. 80% AM 1kHz	Pass
Voltage Dips	EN 61000-4-11	50Hz	0% for 20ms 40% for 200ms 70% for 500ms	Pass
Voltage Interruptions			0% for 5000ms	Pass