EU DECLARATION OF CONFORMITY FOR ATTACHMENT OF 'CE' MARK

DECLARATION OF CONFORMITY FOR ATTACHMENT OF 'UKCA' MARK

We, the manufacturer: BIOQUELL UK Limited.
52 Royce Close,
West Portway
Andover, UK SP10 3TS
Tel: +44 (0) 1264 - 835835

declare under our sole responsibility that the appliance:

CLARUS S

Serial Number as stated on the unit's rating plate

complies with the Essential Requirements of the Machinery Directive 2006/42/EC, EMC Directive 2014/30/EU, the Low Voltage Directive 2014/35/EU and Restriction of Hazardous Substances in Electrical and Electronic Equipment 2011/65/EU by means of the following harmonised standards:

complies with the statutory requirements of the Supply of Machinery (Safety) Regulations 2008, EMC Regulations 2016, Electrical Equipment (Safety) Regulations 2016, The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 by means of the following designated standards:

BS EN 12100:2010
BS EN 61010-1: 2010+A1 2019
BS EN 61326-1:2013
BS EN 50581:2012

A copy of the Technical File for this equipment is available from: Quality Manager

Bioquell SAS, 153 Quai du Raucy, 94380 Bonneuil sur Marne, Paris, FRANCE

Bioquell UK Limited, 52 Royce Close, West Portway, Andover, SP10 3TS, UK

Authorised signatory of manufacturer:

O.R.Cumberlege, Bioquell R&D Group Leader

8th October 2020