## EU DECLARATION OF CONFORMITY FOR ATTACHMENT OF 'CE' MARK



DECLARATION OF CONFORMITY FOR ATTACHMENT OF 'UKCA' MARK

UK CA

We, the manufacturer:

Ecolab Limited. 52 Royce Close, West Portway

Andover, UK SP10 3TS

Tel: +44 (0) 1264 - 835835

declare under our sole responsibility that the appliance:

## **Bioquell ProteQ**

Serial Number as stated on the unit's rating plate

complies with the Essential Requirements of the Machinery Directive 2006/42/EC, Radio Equipment Directive 2014/53/EU, and Restriction of Hazardous Substances in Electrical and Electronic Equipment 2011/65/EU by means of the following harmonised standards:

complies with the statutory requirements of the Supply of Machinery (Safety) Regulations 2008, Radio Equipment regulations 2017, The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 by means of the following designated standards:

Machinery Directive / Regulation: by means of BS EN 12100:2010

Radio Equipment Directive / Regulation by means of:

Article 3.1a	RF Exposure: BS EN 62311
Article 3.1a	BS EN 61010-1:2010+A1:2019
Article 3.1b	BS EN 61326-2-1:2013 using ETSI EN 301 489-3 V2.1.1, & ETSI EN 301
	489-17 V3.1.1 (ETSI EN 301 489-1 V2.1.1)
Article 3.2	EN 300 328 v 2.1.1
	EN 300 330 v 2.1.1

Restriction of Hazardous Substances in Electrical and Electronic Equipment by means of BS EN IEC 63000:2018.

A copy of the Technical File for this equipment is available from: Quality Manager

Ecolab SNC, 153 Quai du Rancy, 94380 Bonneuil-sur-Marne, Paris, FRANCE Ecolab Limited, 52 Royce Close, West Portway, Andover, SP10 3TS, UK

Authorised signatory of manufacturer:

O.R.Cumberlege, Bioquell R&D Group Leader

27th January 2022